

TABLA DE CONVERSIÓN PARA MEDIDOR DE CAUDAL N° 622
Velocidad en m/s

Time in Secs.	1 Rev.	2 Rev.	3 Rev.	5 Rev.	10 Rev.	20 Rev.	30 Rev.	40 Rev.	50 Rev.	60 Rev.	70 Rev.	80 Rev.	90 Rev.	100 Rev.	150 Rev.	200 Rev.	Time in Secs.
40	0.027	0.046	0.064	0.094	0.177	0.344	0.512	0.680	0.847	1.018	1.189	1.356	1.527	1.695	2.542	3.389	40
41	0.027	0.046	0.061	0.091	0.174	0.335	0.500	0.664	0.826	0.994	1.161	1.323	1.490	1.655	2.481	3.307	41
42	0.027	0.043	0.061	0.091	0.171	0.326	0.488	0.649	0.808	0.969	1.134	1.292	1.454	1.615	2.423	3.228	42
43	0.027	0.043	0.061	0.088	0.165	0.320	0.475	0.634	0.789	0.948	1.106	1.262	1.420	1.579	2.368	3.152	43
44	0.027	0.043	0.058	0.085	0.162	0.314	0.466	0.619	0.771	0.927	1.082	1.231	1.387	1.542	2.313	3.078	44
45	0.027	0.043	0.058	0.085	0.158	0.308	0.457	0.607	0.756	0.905	1.058	1.204	1.356	1.509	2.262	3.008	45
46	0.027	0.043	0.058	0.085	0.155	0.302	0.448	0.594	0.741	0.884	1.031	1.180	1.326	1.475	2.213	2.941	46
47	0.024	0.043	0.055	0.082	0.152	0.296	0.439	0.582	0.725	0.866	1.012	1.155	1.298	1.445	2.167	2.880	47
48	0.024	0.043	0.055	0.079	0.149	0.290	0.430	0.570	0.710	0.847	0.991	1.131	1.271	1.414	2.121	2.819	48
49	0.024	0.040	0.055	0.079	0.146	0.283	0.421	0.558	0.695	0.829	0.969	1.106	1.247	1.384	2.076	2.761	49
50	0.024	0.040	0.052	0.079	0.143	0.277	0.411	0.546	0.680	0.814	0.951	1.085	1.222	1.356	2.033	2.710	50
51		0.040	0.052	0.076	0.140	0.274	0.402	0.533	0.668	0.799	0.933	1.064	1.198	1.329	1.993	2.658	51
52		0.040	0.052	0.076	0.140	0.268	0.393	0.524	0.655	0.783	0.914	1.042	1.173	1.305	1.957	2.609	52
53		0.040	0.049	0.073	0.137	0.262	0.387	0.515	0.643	0.768	0.896	1.024	1.152	1.280	1.920	2.560	53
54		0.040	0.048	0.073	0.134	0.259	0.383	0.506	0.631	0.753	0.879	1.006	1.131	1.256	1.894	2.512	54
55		0.040	0.049	0.073	0.131	0.253	0.375	0.497	0.619	0.741	0.863	0.988	1.109	1.234	1.850	2.466	55
56		0.037	0.049	0.070	0.131	0.250	0.369	0.488	0.607	0.728	0.847	0.969	1.091	1.213	1.817	2.423	56
57		0.037	0.049	0.070	0.128	0.244	0.363	0.479	0.597	0.716	0.832	0.951	1.073	1.192	1.786	2.360	57
58		0.037	0.046	0.067	0.125	0.241	0.357	0.469	0.588	0.704	0.817	0.936	1.055	1.170	1.756	2.341	58
59		0.037	0.046	0.067	0.125	0.238	0.351	0.460	0.579	0.692	0.802	0.920	1.036	1.149	1.725	2.301	59
60		0.037	0.046	0.067	0.122	0.235	0.344	0.451	0.570	0.680	0.789	0.905	1.018	1.131	1.693	2.262	60
61		0.037	0.046	0.067	0.119	0.229	0.338	0.445	0.561	0.668	0.777	0.890	1.003	1.113	1.667	2.225	61
62		0.034	0.046	0.064	0.119	0.226	0.332	0.439	0.552	0.658	0.765	0.875	0.988	1.094	1.640	2.188	62
63		0.034	0.043	0.064	0.116	0.223	0.326	0.433	0.543	0.649	0.753	0.860	0.972	1.076	1.615	2.155	63
64		0.034	0.043	0.064	0.116	0.219	0.320	0.427	0.533	0.640	0.741	0.844	0.957	1.061	1.591	2.121	64
65		0.034	0.043	0.061	0.113	0.216	0.314	0.421	0.524	0.631	0.728	0.832	0.942	1.045	1.567	2.088	65
66		0.034	0.043	0.061	0.113	0.213	0.311	0.415	0.515	0.622	0.716	0.820	0.927	1.030	1.542	2.057	66
67		0.034	0.043	0.061	0.110	0.210	0.308	0.408	0.506	0.613	0.707	0.808	0.911	1.015	1.518	2.027	67
68		0.034	0.043	0.061	0.110	0.207	0.305	0.402	0.500	0.604	0.698	0.796	0.899	1.000	1.497	1.996	68
69		0.034	0.040	0.058	0.107	0.204	0.302	0.396	0.494	0.594	0.689	0.783	0.887	0.985	1.475	1.966	69
70		0.034	0.040	0.058	0.107	0.201	0.299	0.390	0.488	0.585	0.680	0.771	0.875	0.969	1.454	1.939	70

Esta tabla se aplica cuando las mediciones son hechas con un medidor suspendido por cable. Cuando las mediciones son hechas con un medidor suspendido por barra, la velocidad tabulada se reduce en un 2%.